

AQS Baden-Württemberg

Proficiency test UKWIR TBT 17 WFD – Tributyltin in surface water

Data assessment sheet

Consensus value

(Only for informative purposes)

Participant:

Date of dispatch: 12 September 2017

GWA Gesellschaft für Wasser-
und Abwasserservice mbH
NL Institut für Wasser- u. Umweltanalytik
An der Ohratalsperre
99885 Luisenthal

Laboratory code: 8


Parameter/Unit	Sample-No.	Assigned Value	Lab result	z	Assessment of the value
Tributyltin ng/l	45	0,5368	0,61	0,5	s
	67	0,3427	0,48	1,6	s
	83	1,368	1,73	1,1	s

According to ISO 13528 (2015) the single results were assessed as follows:

$ z \leq 2.0$	satisfactory
$2.0 < z < 3.0$	questionable
$ z \geq 3.0$	unsatisfactory

Stuttgart, 01 December 2017


Dr.-Ing. Michael Koch
Scientific director AQS


Dr.-Ing. Frank Baumeister
PT manager